



International Conference
2nd International Conference on Innovations, Recent Trends and Challenges
in Mechatronics, Mechanical Engineering and New High-Tech Products
Development
MECAHITECH'10

Bucharest, 23-24 September 2010

An approach for spasticity quantification based on the stretch reflex threshold

João Ferreira, Paulo Rodrigues, Filomena Soares
Industrial Electronics Department, ALGORITMI / University of Minho
Campus de Azurém, 4800-058 Guimaraes, Portugal
fsoares@dei.uminho.pt

José Machado
Mechanical Engineering Department, CT2M / University of Minho
Campus de Azurém, 4800-058 Guimaraes, Portugal
jmachado@dem.uminho.pt

ABSTRACT

Spasticity is a common and complex motor disorder that affects more than 12 million persons in the world. There are several studies on spasticity quantification in the literature but there is still a need for measurement improvements. This paper presents the design of a mechatronic device for spasticity quantification, in joint of ankle, elbow and knees. This approach is based on the velocity dependent of the tonic stretch reflexes. The relevant variables, the measurement range and the adequate measurement systems are selected. The data acquisition system, board and software, are also defined and tested in laboratory. Next step consists of the system validation in clinical environment.

INTRODUCTION

Spasticity is a complex motor disorder due to a supra-spinal inhibition, resulting from a hyper-excitability of the stretch reflex (SR) [1]. Spasticity affects more than 12 million persons in the world [2], and is always seen in patient with upper motor neuron dysfunctions such as cerebral vascular accidents, spinal cord injuries, and multiple sclerosis. The mechanism of spasticity is commonly thought of as an exaggerated SR, which is a velocity-dependent increase in the resistance to the passive movement [3]. The SR threshold is significantly reduced in spastic muscles, and this reduction is correlated to the increase in reflex joint torque [4]. Spasticity in conjugation with excessive muscle tone frequently interferes in the voluntary motor function, causing difficulties in daily activities [3]. Some of the common symptoms are: a change in the recruitment of limb's segments and a severe mal-functioning of the tendons reflex.

The correct quantification of spasticity has been under an extensive study by the scientific community, but there is not yet available a well-accepted standard method for spasticity determination and quantification. The literature presents several methods for spasticity quantification: The Ashworth Scale (AS) and the Ashworth modified version (MAS); Isokinetic device with generator torque; Pendulum Test. None of them is fully accepted, due to various reasons [2], [5], [6]. The AS and the MAS are the common scales in clinical quantifications of spasticity, despite experts agree that both scales may not measure the characteristics that distinguish spasticity from other tonus disorders. Although the scale is useful in determining the amount of resistance felt in the passive displacement of the limbs, it



International Conference
2nd International Conference on Innovations, Recent Trends and Challenges
in Mechatronics, Mechanical Engineering and New High-Tech Products
Development
MECAHITECH'10

Bucharest, 23-24 September 2010

does not quantify the dependence to velocity, which is the feature that differentiates spasticity [5]. This scale has a low reproductive rate, a lack of validation in all muscle groups, usually affected by spasticity [1]. The approach of traditional measures is based on the phase and magnitude of the tonic SR and the resistance to passive stretch. Nevertheless, this measure is not correlated to the clinical impression of the spasticity degree, inability to differentiate the mechanical stiffness from the reflective stiffness and the implementation of the device is still complex; also, the measurement does not meet the criteria of the known theory. Hence, there is still a need for a device that meets these requirements [1], [5], [6].

The key issue is to determine which variables are necessary to correctly quantify this disorder. A correct measure to quantify spasticity must follow the physiological mechanisms related to the stand-up position control and the movement in healthy individuals and/or must detect possible deficiencies in any of these mechanisms that lead to motor disorders. For the acceptance of the method, their approach must be in accordance with a standard spasticity definition [5], [6]. The works presented in [7] defined spasticity as: "a motor disorder characterized by a velocity-dependent increase in tonic stretch reflexes ("muscle tonus") with exaggerated tendon jerk, resulting in hyper excitability of the stretch reflexes, as one of the component of the upper motor neuron syndrome". This definition is still accepted nowadays; it includes some important aspects: it refers to spasticity as a symptom, as a disorder in the somatic mobility, related to the high tonic component of the SR; it is due to the spinal reflex; it is one of the symptoms of the upper motoneuron syndrome; the tonic stretch is associated to the exaggerated tendon jerks, and reflects the physical component of the SR; the reflex of the tone stretch is the basis of the tonus; it is referred that the excess of the reflex depends on the stretch velocity [8]. This last statement is the key issue for spasticity quantification [2].

This paper presents the study on the design of a mechatronic device for quantification of all levels of spasticity, in joint of ankle, elbow or knees. This approach is based on the velocity dependent increase in the tonic SR, according to the criteria of spasticity definition proposed by Lance [7], to establish the relationship, for clinical evaluation, between all levels of spasticity.

The article is divided in four sections: section one introduces the spasticity concept and the problem formulation; section two presents the proposed approach used for spasticity quantification; section three discusses the obtained results and explains related ongoing works and, finally, section four resumes the conclusions.

PROPOSED APPROACH

The SR is an involuntary contraction elicited by a brief stimulus to muscle receptors. If the arm and muscle are immobilized the result will be a measurable change in the tension of the tendon [4].

The objective of this work is to develop a device for the quantification of all levels of spasticity, which can be accepted by the scientific community. It aims to develop universal device that allows the evaluation for the joint of ankle, knees or elbow. In this approach the method is focused on the velocity dependent increasing in the Tonic SR, according to the criteria of standard definition of Spasticity, proposed by Lance [7]. This definition suggested that the Stretch Reflex Threshold (SRT) depends on the velocity of stretch.

It is determined the angle of biomechanics range and the angular velocity when an increase in electromyography activity occurs, for further data processing, by a program developed especially for this propose. Most of the daily life activity requires joint angles of

45° for ankle, 140° to full flexion to the knee [9] and 120° for elbow [10]. These ranges should be considered to ensure the assessment of movement in the whole biomechanical range, recruited in daily life. The study proposed in [2], where are studied patients with stroke, there is detected an increase of activity of the EMG (electromyography) signal in Biceps Brachia, due to stretching the elbow joint, at angular velocities of 51°/s, 161°/s and 430°/s; this demonstrates the dependence of velocity of SR, in muscle affected by spasticity.

To determine the SR Threshold we propose three measurements: EMG signal activity in the muscle; angular velocity of passive muscle stretch and the joint angle position.

Table 1 resumes the parameters and the measuring ranges for the proposed equipment.

Table 1 Variable Measuring Range in the Proposed Approach

| Parameters | Measuring range |
|------------------|-----------------|
| angle | 180° |
| electromyography | 20-500 Hz |
| Angular velocity | 500°/s |

Experimental setup

The experimental set-up protocol (Figure 1) consists of the following steps.

The patient sits on a chair and the electromyography sensors are fixed on a motor point of a biceps brachial and the EMG signal at rest (tone) is recorded. To record the biomechanical angle, biomechanical range and compute the angular velocity a goniometer is fixed to the arm of the individual. The axis of rotation of the goniometer is placed in correspondence to the elbow joint. The arm of goniometer is aligned and fixed to the arm and forearm of the patients. The patient is placed in the ideal joint position. The initial angle position and biomechanical range of joint are recorded and constantly monitored, to ensure maximum repeatability and maximal Reflex response of the Biceps Brachial. In each stretch the initial angle joint is checked and only starts a new stretch if this condition is confirmed. The beginning of each stretch only occurs if the EMG signal, in the initial position, corresponds to the signal captured at rest, (the muscle tone).




Fig. 1: Proposed system for spasticity quantification.

Experimental procedure

In order to determine the dynamic SRT (DSRT), the muscle is stretched manually at different velocities. DSRT is defined as the joint angle and the corresponding velocity value, when the EMG signal amplitude also increases with the velocity, above the threshold corresponding with the EMG signal amplitude at rest. The Tonic SRT (TSRT) represents a specific value of DSRT when velocity equals to zero (at rest). In this approach, the DSRT and the TSRT are expressed in velocity and angular coordinates values. The SRT can be expressed as a specific point in the range of joint angle. Thus the SRT can be related to the body frame of reference.

In Figure 2 the DSRT is plotted on a coordinate system, in two dimensions, angular velocity versus angle. Regression analyses can be used in order to compute TSRT value, by extrapolating the regression line through the points of DSRT. When the regression line crosses the axes corresponding at velocity value equal to zero, it corresponds to the coordinate angle at rest (TSRT).




Fig: 2 Diagram Velocity (°/s) versus angle (°). Adapted from [3].

Figure 3 shows the relation between the range of regulation of TSRT and biomechanical range of the joint [1]. The grey areas to the right of diagonal lines indicate the areas where spastic muscle is active [2], [13]. Previous studies with animals and humans suggest that in healthy individuals the range of regulation λ (Figure 3) beyond biomechanical range of joint (θ^- , θ^+), where θ^- is the start position of the passive stretch and θ^+ is the end position of stretch. In sick individuals the range of regulation λ lies within the biomechanical range of joint.




Fig: 3 Relation between range of regulation of STRT and biomechanical range of joint. Adapted from [13].

EMG Signal detection

Electromyography records the electrical activity of muscle, and it is a powerful tool in the analysis of human muscular system. When the muscles are active they produce an electric current generally proportional to muscle activity. EMG studies the muscle function through the interpretation of bioelectric signal produced by the muscle.

To measure the EMG signal, surface or needle electrodes are used, depending on the muscle type, superficial or deep. Passive surface electrodes have no amplification in the electrode; active surface electrodes have a signal preamplification system before being sent to the conditioner, which enable a noise reduction.

In this approach, passive surface Ag/AgCL electrodes are used and they do not cause pain to the patient. SENIAM [11] recommends the use of electrodes Ag/AgCl, together with a conductive gel to reduce signal noise by ensuring a better contact between the electrode and the skin. The signal muscle when measured using surface electrodes has amplitude to 5mV. The frequency range of the EMG signal for the correct analysis is limited to 20Hz and 500Hz, since frequencies below 20Hz tend to fluctuate and to be unstable.

The electrodes have a bipolar configuration, enabling a high rate of common mode rejection, and easily eliminating/reducing signal noise. They should be placed in a 20 mm distance from each other. A surface cleaning gel should also be used in order to reduce the impedance skin/electrode.

The raw EMG signal is detected, amplified and sampled with an analog to digital converter (ADC), after filtering with an anti-aliasing filter. The detection algorithm is implemented in software in LabVIEW Software, Laboratory Virtual Instrument Engineering Workbench, from National Instrument so that the user can monitor the procedure and the results. The EMG signal increase detection is calculated by the method proposed in [12].

Precise detection of on-off timing of human skeletal muscle during movement, based on surface electromyography (sEMG), is an important issue in the analysis of the motor system. The results depend on the choice of threshold. The goal is to perform signal estimation from noise contaminated EMG signal. The common method for resolving motor-

related events from EMG signals consists of visual impression by trained observers. The “single-threshold method”, which compares the EMG signal to fixed threshold, is generally unsatisfactory. Double threshold detection method proposed by [12] is better than single-threshold because it yields higher detection probability and higher sensitivity. Double-detection allows the user to adapt the link between false alarm and detection probability with a higher degree of freedom than the single-threshold [13], [14].




Fig.4. Detection results for Vastus Lateralis muscle group; (a) original raw EMG signal; (b) results of proposed method on-off detection [12]

RESULTS AND DISCUSSION

The more important data for spasticity quantification are, as already indicated, angular velocity, joint angle and electromyography signal activity in the muscle. A precise data acquisition related to these electrical signals can assure us that our developed mechatronic device is working as expected and, also, that is able to allow us to quantify spasticity according the previous presented methods. Figures 5 and 6 show how successful was the data acquisition done with our developed device. Fig. 5 shows the acquisition related with joint angle and angular velocity and Fig. 6 shows the acquisition related with electromyography signal activity in the muscle. The tests were performed simulating a real exam with variable angular velocities, resulting in the EMG signal and angle for every instant of tests.




Fig.5. Angle (left) and velocity (right) graphics




Fig.6. Original raw EMG signal

Concerning the developed system, presented in this paper, we can assure that it is able to solve the problem of spasticity quantification.

The most important tasks for this device are the correct data acquisition related to angular velocity, joint angle and electromyography signal activity in the muscle.

The device is, still under, development determination of the double threshold detection following the method proposed by [12], and to apply the developed device to measure spasticity in clinic environment. The informal consent is just to be authorized and signed.

CONCLUSIONS AND FUTURE WORK

This paper presents the study and design of a device for spasticity quantification, to be used in joint of ankle, elbow and knees. This approach is based on Lance's work [7] that states the velocity dependence increase in the TSR. To determine the SRT, three measurements were proposed: EMG signal activity in the muscle; angular velocity of passive muscle stretch and the joint angle position.

The sensor devices were chosen taking into account the system variables and the corresponding measurement range. In particular, the electromyography signal has a measuring range of 20-500 Hz; the angle, 180°; and the angular velocity, 500°/s.

The data acquisition system, board and software, selected are from National Instrument due to its functionalities. The USB interface of the board enables system portability, an important requirement of the proposed device.

The system prototype was validated in laboratorial environment.

Next step consists of the implementation, testing and validation of the proposed mechatronic device in clinical environment, by medical staff, being the informal consent with a public Portuguese hospital about to be signed.

REFERENCES

- [1] Le Carvozin, P., X. Hemot, O. Bartier, G. Carrault, F. Chagneau, P. Callien, H. Allain, P. Rochcongar, "Evaluation of pendulum testing of spasticity" in *Annales de la Réadaptation et de Médecine Physique*, Mars. 2002, 45, pp. 510-516.
- [2] Calota, A., A.G. Feldman, M.F. Levin "Spasticity measurement based on tonic stretch reflex threshold in stroke using a portable device" In *Clinical Neurophysiology*. 2008, 119, pp. 2329-2337.



International Conference
2nd International Conference on Innovations, Recent Trends and Challenges
in Mechatronics, Mechanical Engineering and New High-Tech Products
Development
MECAHITECH'10

Bucharest, 23-24 September 2010

- [3] Chen, J.J, Y.N. Wu, S.C Huang, H.M. Lee and Y.L. Wangs “The Use of a Portable Muscle Tone Measurement Device to Measure the Effects of Botulinum Toxin Type A on Elbow Flexor Spasticity”. Arch Phys Med Rehabil,. Vol. 86, Agu. 2005.
- [4] Chardon M., L. Nina, Surech and w.z. Ryner. “A New Method For Reflex Threshold Estimation in Spastic Muscles” Presented at the conference of the IEEE EMBS. Sep. 2009.
- [5] Levin, M.F., “On the nature and measurement of spasticity” in Clinical Neurophysiology, 2005. 116; 1754-1755
- [6] Bernhard, J L. “Towards a Clinical measurement of Spasticity” In Physiotherapy. 1995 81.
- [7] Lance, J.W. “Spasticity: Disorder of Motor Control” Chicago, Year Book Medical, 1980, pp. 485-500.
- [8] Decq, P., P. Pilipetti, J.P. Lefaucheur “Evaluation of Spasticity in Adults” Operative Technique in Neurosurgery. 2005, PP. 100-108.
- [9] Febiger, L. “Kinesiology and Applied Anatomy (7th ed.)” Guanabara Koogans. A. 1989, PP. 146-170.
- [10] Ennis, O., D. C. Miller, P. Kelly,. “Fracture of the adult elbow” in Current Orthopaedics. 2008, 22; pp. 111-131.
- [11] Hermens, H. J., B.Freriks, C. Disselhorst-Klug, G. Rau “Development of recommendations for SEMG sensors and sensor placement procedures” In Journal of Electromyography and Kinesiology. 2000, 10; pp. 361–374.
- [12] Lanyi, X., A. Adler “An improved method for muscle activation detection during gait” Canadian Conference of Electrical and Computer Engineering 2004.
- [13] Jobin, A. and M.E.Levin. “Regulation of stretch reflex threshold in elbow flexors in children with cerebral palsy: a new measure of spasticity” Developmental Medicine & Child Neurology. 2000, 42; pp. 531-540.
- [14] Reaz, M.B.I., M.S.S Hussain and F.Mohd-yasin “Techniques of EMG signal analysis: detection, processing, classification and aplication” In Biol. Proced. Online. 2006; 8(1):11-35.